

PrecisionBind Human Interleukin 8 (IL-8) ELISA

REF: KB1070

Ver 1.0






RUO

NIBSC Calibrated Assay

*the standards used in this kit are calibrated against an international standard from the National Institute of Biological Standards and Control (NIBSC), Potters Bar, Hertfordshire EN6 3QG, UK.

1 ng of supplied standard equals 580 U of 93/730 NIBSC-standard. Please note that the calibration is lot specific.

ELISA for Accurate Quantitation of Human IL-8 from Cell Culture Supernatant, Serum, Plasma, or Other Bodily Fluids

RUO	For Research Use Only	REF	Catalog Number
	Store At	LOT	Batch Code
	Manufactured By		Biological Risk
	Expiry Date		Consult Operating Instructions

For Research Purposes Only. Not for use in diagnostic or therapeutic procedures. Purchase does not include or carry the right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Krishgen Biosystems Private Limited is strictly prohibited.

REF KB1070

 96 tests

Krishgen Biosystems Private Limited

For US/Europe Customers: toll free +1(888)-970-0827 | tel +1(562)-568-5005

For Asia/India Customers: tel +91(22)-49198700

Email: sales1@krishgen.com | <http://www.krishgen.biz> / www.krishgenbio.com

PrecisionBind Human Interleukin 8 (IL-8) ELISA

Introduction:

IL-8 is a potent lymphoid cell growth factor that stimulates the growth and survivability of certain B cells and T cells. IL-8 plays a role in host defense, acute phase reactions, immune response, and hematopoiesis. IL-8 is expressed by T cells, B cells, monocytes, fibroblasts, hepatocytes, endothelial cells, and keratinocytes. Recombinant Human IL-8 is a 20.9 kDa protein containing 184 amino acids.

Long Name:

Interleukin 8

Entrez Gene IDs:

3576 (Human); 396880 (Porcine); 403850 (Canine); 493836 (Feline)

Alternate Names:

3-10C; AMCF-I; C-X-C motif chemokine 8; CXCL8; CXCL8SCYB8; Emotakin; GCP1; GCP-1TSG-1; IL8; IL-8; interleukin 8; K60; LAI; LECT; LUCT; LYNAP; MDNCF; MDNCFb-ENAP; member 8; MONAP; MONAPGCP1; NAF; NAP1; NAP-1NAP1; NCF; Neutrophil-activating protein 1; Protein 3-10C; T cell chemotactic factor; T-cell chemotactic factor; TCF; TSG1

Intended Use:

PrecisionBind Human Interleukin 8 (IL-8) ELISA is specifically designed for the accurate quantitation of Human IL-8 from cell culture supernatant, serum, plasma or other bodily fluids. It is ready-to-use, accurate, and sensitive.

Principle:

This assay is based on the Sandwich ELISA procedure. Samples containing human IL-8 react with already coated affinity purified capture anti-Human IL-8 antibody and bind to them. Plates are washed with wash buffer to remove unbound reactants. Biotinylated Anti-Human IL-8 is added leading to formation of a sandwich complex of solid phase antibody-Human IL-8-biotin labeled antibody. The wells are washed to remove any unbound reactants as per the wash procedure. Streptavidin:HRP conjugate is added which binds to Biotinylated Anti-Human IL-8 complex. The wells are washed to remove any unbound reactants as per the wash procedure. The substrate 3, 3',5, 5' Tetra Methyl Benzidine (TMB) is then reacted. The amount of hydrolyzed substrate is read on a microtiter plate reader at 450 nm and it is directly proportional to the concentration of Human IL-8 present in the samples.

Materials Provided:

1. Anti-Human IL-8 Coated Microtiter Plate (12 x 8 wells) – 1 no.
2. Recombinant Human IL-8 Standard (lyophilized) – 2 vials
3. Anti-Human IL-8 Biotin Conjugated Detection Antibody - 1 vial
4. Streptavidin:HRP Conjugate (concentrated) - 1 vial
5. Streptavidin:HRP Diluent - 12 ml
6. Assay Diluent - 50 ml
7. (20X) Wash Buffer - 25 ml
8. TMB Substrate - 12 ml
9. Stop Solution - 12 ml
10. Instruction Manual

Materials to be provided by the End-User:

1. Microplate Reader able to measure absorbance at 450 nm.
2. Adjustable pipettes to measure volumes ranging from 50 ul to 1000 ul.
3. Deionized (DI) water.
4. Wash bottle or automated microplate washer.
5. Graph paper or software for data analysis.
6. Tubes to prepare standard/sample dilutions.
7. Timer.
8. Absorbent paper.

PrecisionBind Human Interleukin 8 (IL-8) ELISA

Storage Information:

1. Store main kit components at 2-8°C.
2. Store recombinant Standard at 2-8°C. Upon reconstitution, aliquot recombinant protein into polypropylene vials and store at -20°C as per assay requirements. Do not freeze thaw for more than two times.
3. Before using, bring all components to room temperature (18-25°C). Upon assay completion return all components to appropriate storage conditions.

Health Hazard Warnings:

1. Reagents that contain preservatives may be harmful if ingested, inhaled or absorbed through the skin. Refer to the MSDS online for details.
2. To reduce the likelihood of blood-borne transmission of infectious agents, handle all serum and/or plasma in accordance with NCCLS regulations.

Specimen Collection and Handling:

Specimens should be clear and non-hemolyzed. Samples should be run at a number of dilutions to ensure accurate quantitation.

Cell Culture Supernatant: If necessary, centrifuge to remove debris prior to analysis. Samples can be stored at temperature <-20°C. Avoid repeated freeze/thaw cycles.

Serum: Use a serum separator tube and allow clotting for 30 minutes, then centrifuge for 10 minutes at 1000 x g. Remove serum layer and assay immediately or store serum samples at temperature <-20°C. Avoid repeated freeze/thaw cycles.

Plasma: Collect blood sample in a citrate, heparin or EDTA containing tube. Centrifuge for 10 minutes at 1000 x g within 30 minutes of collection. Assay immediately or store plasma samples at temperature <-20°C. Avoid repeated freeze/thaw cycles.

Reagent Preparation:

Please refer to lot-specific instructions for preparation of the reagents mentioned in the Reagent Preparation Sheet. Note each reagent sheet is specific for a particular Lot only and is not to be interchanged amongst different lots.

Procedural Notes:

1. In order to achieve good assay reproducibility and sensitivity, proper washing of the plates to remove excess un-reacted reagents is essential.
2. High Dose Hook Effect may be observed in samples with very high concentrations of Human IL-8. High Dose Hook Effect is due to excess of antibody for very high concentrations of Human IL-8 present in the sample.
3. Human IL-8 concentration of the undiluted sample is less than the diluted sample, this may be indicative of the Hook Effect.
4. Avoid assay of Samples containing sodium azide (NaN₃), as it could destroy the HRP activity resulting in under-estimation of the amount of Human IL-8.
5. It is recommended that all Standards and Samples be assayed in duplicates or triplicates.
6. Maintain a repetitive timing sequence from well to well for all the steps to ensure that the incubation timings are same for each well.
7. If the Substrate has a distinct blue color prior to use it may have been contaminated and use of such substrate can lead to compromisation of the sensitivity of the assay.
8. The plates should be read within 30 minutes after adding the Stop Solution.
9. Make a work list in order to identify the location of Standards and Samples.

Assay Procedure:

1. It is strongly recommended that all Standards and Samples be run in duplicates or triplicates. A standard curve is required for each assay

PrecisionBind Human Interleukin 8 (IL-8) ELISA

2. Add 100 ul of prepared **Standards** and **Samples** to the respective wells.
3. Add 50 ul of **diluted Biotinylated Detection Antibody** solution to each well. Mix well. Seal plate and incubate for 2 hours at 37°C.
4. Aspirate and wash plate 4 times with **Wash Buffer (1X)** and blot residual buffer by firmly tapping plate upside down on absorbent paper. Wipe of any liquid from the bottom outside of the microtiter wells as any residue can interfere in the reading step. All the washes should be performed similarly.
5. Add 100 ul of **diluted Streptavidin:HRP** solution to each well. Seal plate and incubate for 30 minutes at 37°C.
6. Wash plate 4 times with **Wash Buffer (1X)** as in step 4.
7. Add 100 ul of **TMB Substrate** solution and incubate in the dark for 30 minutes at 37°C .Positive wells should turn bluish in color. It is not necessary to seal the plate during this step.
8. Stop reaction by adding 100 ul of **Stop Solution** to each well. Positive wells should turn from blue to yellow.
9. Read the absorbance at 450 nm with a microplate within 10-15 minutes after addition of Stop solution.

Calculation of Results:

Determine the mean absorbance for each set of duplicate or triplicate standards and samples. Subtract the mean absorbance of the zero standards (background) from each well.


Plot the standard curve on standard graph paper, with cytokine concentration on the x-axis and absorbance on the y-axis. Draw the best fit straight line through the standard points. To determine the unknown cytokine concentrations, find the unknowns mean absorbance value on the y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the x-axis and read the cytokine concentration. If samples were diluted, multiply by the appropriate dilution factor.

Computer based curve-fitting software may be preferred. Software which is able to generate a cubic spline curve-fit or a 4-PL (2nd order) is best recommended for automated results.

Typical Data (representative only)

Standard Concentration (pg/ml)	Abs A	Abs B	Mean Abs	Interpolated Concentration	% Interpolated Concentration against Actual Concentration	% STD Deviation	%CV
0	0.096	0.097	0.096	--	--	0.1	0.7
15.62	0.186	0.178	0.182	18.4	117.8	0.6	3.2
31.25	0.236	0.248	0.242	35.0	112.1	0.9	3.6
62.5	0.426	0.401	0.414	73.3	117.3	1.8	4.3
125	0.640	0.682	0.661	122.0	97.6	3.0	4.5
250	1.287	1.213	1.250	240.7	96.3	5.2	4.2
500	2.159	2.273	2.216	515.4	103.1	8.0	3.6
1000	3.069	2.917	2.993	986.8	98.7	10.7	3.6

Typical Graph (representative only)



PrecisionBind Human Interleukin 8 (IL-8) ELISA

Performance Characteristics:

Please note that this validation is performed in our laboratory and will not necessarily be duplicated in your laboratory. This data has been generated to enable the user to get a preview of the assay and the characteristics of the kit and is generic in nature. We recommend that the user performs at the minimum; the spike and recovery assay and the dilutional linearity assay to assure quality results. For a more comprehensive validation, the user may run the protocols as suggested by us herein below to develop the parameters for quality control to be used with the kit.

Sensitivity:

Limit Of Detection: It is defined as the lowest detectable concentration corresponding to a signal of Mean of '0' standard plus 2*SD. 10 replicates of '0' standards were evaluated and the LOD is **~0.2 pg/ml**.

Limit of Quantitation (LOQ): It is defined as the lowest concentration of an analyte that can be measured with acceptable precision and accuracy, 10 replicates of '0' standards were evaluated and the LOQ is **~0.6 pg/ml**.

IC₅₀: The half-maximal inhibitory concentration (IC₅₀) in a sandwich ELISA measures the concentration of an inhibitor (such as a drug, molecule, or antibody) required to reduce the binding of a target antigen to the capture/detection antibody pair by 50%. The IC₅₀ for PrecisionBind Human IL-8 ELISA is **~224 pg/ml**.

Specificity:

The antibodies used in the kit for capture and detection are monoclonal antibodies specific for Human IL-8. The standards used in this kit are **calibrated against an international standard from the National Institute of Biological Standards and Control (NIBSC)**, Potters Bar, Hertfordshire EN6 3QG, UK. 1 ng of supplied standard equals 580 U of 93/730 NIBSC-standard. Please note that the calibration is lot specific.

Cross-Reactivity:

This assay recognizes natural and recombinant human IL-8. The markers listed below were prepared at 250 pg/ml in Assay Diluent and assayed for cross-reactivity. No significant cross-reactivity or interference was observed.

Recombinant Human:

ANG	AR	CNTF	β-ECGF	EGF	Epo	FGF acidic	FGF basic	FGF-4
FGF-5	FGF-6	G-CSF	GM-CSF	GROα	GROβ	GROγ	gp130	HB-EGF
HGF	I-309	IFN-γ	IGF-I	IGF-II	IL-1α	IL-1β	IL-1ra	IL-1 RI
IL-1 RII	IL-2	IL-2 Rα	IL-3	IL-3 Rα	IL-4 R	IL-5	IL-5 Rβ	IL-6
IL-6 R	IL-7	IL-9	IL-10	IL-11	IL-12	IL-13	IP-10	KGF
LAP (TGF-β1)	LIF	M-CSF	MCP-1	MCP-2	MCP-3	MIP-1α	MIP-1β	β-NGF
OSM	PD-ECGF		PDGF-AA	PDGF-AB	PDGF-BB	PF-4		PTN
RANTES	SCF	SLPI	TGF-α	TGF-β1	TGF-β3	TGF-β RII		TNF-α
TNF-β	TNF RI	TNF RII	VEGF					

Recombinant Mouse:

GM-CSF IL-1α IL-1β IL-3 IL-4 IL-5 IL-5 Rα IL-6 IL-7 IL-9 IL-10 IL-13 KC LIF MIP-1α MIP-1β SCF TNF-α

Recombinant Amphibian:

TGF-β5

Natural Proteins:

bovine FGF acidic bovine FGF basic human PDGF porcine PDGF human TGF-β1 porcine TGF-β1 porcine TGF-β2

Assay Range:

15.62 pg/ml to 1000 pg/ml.

Parallelism and Matrix Effect

Sample Dilution factor – Human Serum, Human Plasma and Human CSF samples have been tested. Sample dilution Factor for all three matrices is 1:50 dilution.

Neat Human Serum, Human Plasma and Human CSF were spiked with 250 pg/ml Human IL-8 and ELISA assay was run.

Sample	Mean OD450	Interpolated Concentration	% Recovery
Neat CSF samples	2.971	528.2	Out of range
Neat Plasma	3.175	576.9	230.8
Neat Human Serum	3.818	735.1	294.0

PrecisionBind Human Interleukin 8 (IL-8) ELISA

A) Serum:

Dilution	Expected Standard Concentration (pg/ml)	Mean Absorbance	Interpolated Concentration (pg/ml)	% Recovery	% Deviation
1:100 dilution	500	2.352	436.9	87.4	114.4
1:200 dilution	250	1.598	246.5	98.6	101.4
1:400 dilution	125	0.913	100.6	80.5	124.3
1:800 dilution	62.5	0.731	58.2	93.1	107.4
1:1600 dilution	31.25	0.555	28.7	91.8	108.9
1:3200 dilution	15.6	0.278	4.8	30.7	325.8
1:6400 dilution	7.8	0.200	--	0.0	--

B) Plasma:

Dilution	Expected Standard Concentration (pg/ml)	Mean Absorbance	Interpolated Concentration (pg/ml)	% Recovery	% Deviation
1:100 dilution	500	2.569	434.9	87.0	115.0
1:200 dilution	250	1.652	236.5	94.6	105.7
1:400 dilution	125	1.156	140.1	112.1	89.2
1:800 dilution	62.5	0.810	79.3	126.9	78.8
1:1600 dilution	31.25	0.567	41.3	132.0	75.8
1:3200 dilution	15.6	0.313	8.1	52.0	192.5
1:6400 dilution	7.8	0.241	1.3	16.2	616.8

C) Cerebrospinal Fluid (CSF):

Dilution	Expected Standard Concentration (pg/ml)	Mean Absorbance	Interpolated Concentration (pg/ml)	% Recovery	% Deviation
1:100 dilution	500	2.610	444.3	88.9	112.5
1:200 dilution	250	1.674	240.9	96.4	103.8
1:400 dilution	125	1.178	144.1	115.3	86.7
1:800 dilution	62.5	0.832	83.0	132.8	75.3
1:1600 dilution	31.25	0.590	44.6	142.9	70.0
1:3200 dilution	15.6	0.343	11.5	73.5	136.2
1:6400 dilution	7.8	0.268	3.6	45.8	218.8

Results:

- i) Parallelism is maintained across the 1:100 to 1:1600 dilutions.
- ii) % Recovery for most dilutions falls within the acceptable range of 80% - 120%.
- iii) No significant matrix effect observed at higher dilutions.
- iv) The PrecisionBind Human IL-8 ELISA kit was tested for matrix effect on human serum, plasma, CSF and physiological buffer 7.4 to mimic tear fluid samples.

Precision:

Intra-Assay: CV<10%
Inter-Assay: CV<12%

Linearity:

The linearity of the kit was assayed by testing samples spiked with appropriate concentration of Human IL-8 and their serial dilutions.
The results were demonstrated by the percentage of calculated concentration to the expected.

Sample	1:2	1:4	1:8
serum (n=5)	84-107%	87-108%	82-112%
EDTA plasma (n=5)	83-102%	83-115%	83-118%
heparin plasma (n=5)	83-99%	80-95%	82-93%

PrecisionBind Human Interleukin 8 (IL-8) ELISA

Safety Precautions:

- **This kit is for research use only.** Follow the working instructions carefully.
- The expiration dates stated on the kit are to be observed. The same relates to the stability stated for reagents
- Do not use or mix reagents from different lots.
- Do not use reagents from other manufacturers.
- Avoid time shift during pipetting of reagents.
- All reagents should be kept in the original shipping container.
- Some of the reagents contain small amount of sodium azide (<0.1% w/v) as preservative. They must not be swallowed or allowed to come into contact with skin or mucosa.
- Source materials maybe derived from **Human body fluids** or organs used in the preparation of this kit were tested and found negative for HBsAg and HIV as well as for HCV antibodies. However, no known test guarantees the absence of such viral agents. Therefore, handle all components and all patient samples as if potentially hazardous.
- Since the kit contains potentially hazardous materials, the following precautions should be observed
 - Do not smoke, eat or drink while handling kit material
 - Always use protective gloves
 - Never pipette material by mouth
 - Wipe up spills promptly, washing the affected surface thoroughly with a decontaminant.
- In any case GLP should be applied with all general and individual regulations to the use of this kit.



References:

Serum and tumor microenvironment IL-8 values in different stages of colorectal cancer.

M Bălășoiu, AT Bălășoiu, SȘ Mogoantă... - Rom J Morphol ..., 2014 - researchgate.net... IL-8 was determined using ELISA Sandwich technique, with the kit Human IL-8 ELISA (Krishgen Biosystem Private Limited), in the presence of the concentration standards. Double sample ...

Role of interleukin-8 in the assessment of innate immune response in respiratory acute infections in breastfed infants.

M HILA - Acta Medica Transilvanica, 2016 - amtsibiu.ro ... In order to determine the serum level of interleukin-8 (IL-8), I used the Human IL-8 ELISA (Cat.No: KB1070) Krishgen BioSystem's research kit. Being a kit dedicated only for dosages ...

Yashada bhasma (Zinc calx) and Tankana (Borax) inhibit Propionibacterium acne and suppresses acne induced inflammation in vitro.

R Sandeep Varma, S Shamsia... - ... Journal of Cosmetic ..., 2014 - Wiley Online Library ... TNF- α and IL-8 screening kits were purchased from Krishgen Biosystems Private Limited (Mumbai, India). All other chemicals used in the experiments were of molecular biology grade. ...

The biochemical role of different classes of cytokine in cancer.

E Simionică, O Gheorghită - Scientific Collection" InterConf", 2021 - ibn.idsi.md.... Interleukin 8 values were measured using the Sandwich ELISA technique using the Human IL-8 ELISA kit produced by Krishgen BioSystem in Spain, in the presence of standard ...

Asiatic acid inhibits intracellular Shigella flexneri growth by inducing antimicrobial peptide gene expression.

P Maitra, P Basak, K Okamoto, S Miyoshi... - Journal of Applied ..., 2023 - academic.oup.com ... ELISA was performed to check IL-8... IL-8 secreted into the cultured media of infected and treated HT29 cells using Cathelicidin ELISA KIT (MyBioSource) (#MBS720523), GENLISA ELISA ...

Role of STAT signaling and autocrine action of chemokines during H₂O₂ induced HTR-8/SVneo trophoblastic cells invasion.

P Banerjee, A Malik, SS Malhotra... - Journal of Cellular ..., 2019 - Wiley Online Library ... IL-8 at protein level was also confirmed by ELISA using Human IL-8 ELISA Kit (Krishgen Biosystems Private Limited) and Human ...determine the concentration of IL-8 and MIP-1 β in ELISA. Cells were ...

SCHEMATIC ASSAY PROCEDURE

1. Remove all components, 30 minutes before adding into the assay plate.



2. Avoid repeated cool-thaw of the components as there will be a loss of activity and this can affect the results.



3. Pipette **100 ul Standards/ Samples** into respective Standard wells.

4. Pipette **50 ul diluted Biotinylated Detection antibody** to all wells. Mix well.

5. Cover plate and incubate for at 37°C.

6. Aspirate and wash wells 4 times with **Wash Buffer (1X)**.

7. Pipette **100 ul diluted Streptavidin:HRP** to all wells.

8. Cover plate and incubate for at 37°C.

9. Aspirate and wash wells 4 times with **Wash Buffer (1X)**.

10. Pipette **100 ul TMB Substrate** into each wells.

11. Cover plate and incubate for at 37°C.

12. Pipette **100 ul Stop Solution** into each well.

13. Read absorbance at 450 nm with a microplate reader within of stopping reaction.

LIMITED WARRANTY

Krishgen Biosystems Private Limited does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the Products; against defects in products or components not manufactured by Krishgen Biosystems Private Limited, or against damages resulting from such non-Krishgen Biosystems Private Limited made products or components. Krishgen Biosystems Private Limited passes on to customer the warranty it received (if any) from the maker thereof of such non Krishgen made products or components. This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Krishgen Biosystems Private Limited.

THIS WARRANTY IS EXCLUSIVE. The sole and exclusive obligation of Krishgen Biosystems Private Limited shall be to repair or replace the defective Products in the manner and for the period provided above. Krishgen Biosystems Private Limited shall not have any other obligation with respect to the Products or any part thereof, whether based on contract, tort, and strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall Krishgen Biosystems Private Limited be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of Krishgen Biosystems Private Limited with respect to the Products. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

Krishgen Biosystems Private Limited. 2026














THANK YOU FOR USING A KRISHGEN PRODUCT!

KRISHGEN BIOSYSTEMS PRIVATE LIMITED®, DHARMAPLEX®, GENBULK®, GENLISA®, KRISHZYME®, KRISHGEN®, KRIBIOLISA®, KRISHPLEX®, TITANIUM®, QUALICHEK®, PRECISIONBIND® are registered trademarks of KRISHGEN BIOSYSTEMS PRIVATE LIMITED.

©KRISHGEN BIOSYSTEMS PRIVATE LIMITED. ALL RIGHTS RESERVED.

KRISHGEN BIOSYSTEMS PRIVATE LIMITED | OUR REAGENTS | YOUR RESEARCH |

SYMBOLS KEY

	Anti-Human IL-8 Coated Microtiter Plate (12x8 wells)
	Recombinant Human IL-8 Standard, Lyophilized
	Anti-Human IL-8 Biotin Conjugated Detection Antibody
	Concentrated Streptavidin Horseradish Peroxidase
	Streptavidin:HRP Diluent
	Assay Diluent
	(20X) Wash Buffer
	TMB Substrate
	Stop Solution
	Consult Instructions for Use
	Catalogue Number
	Expiration Date
	Storage Temperature