






KRIBIOLISA™ Natalizumab (TYSABRI™) ELISA

REF : KBI1020

Ver 1.1

RUO

Enzyme Immunoassay for the Quantitative Determination of
Natalizumab in human serum and plasma.

| | | | |
|---|------------------|---|--------------------------------|
| RUO | For Research Use | REF | Catalog Number |
|  | Store At | LOT | Batch Code |
|  | Manufactured By |  | Biological Risk |
|  | Expiry Date |  | Consult Operating Instructions |

For Research Use Only. Purchase does not include or carry the right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of KRISHGEN BioSystems is strictly prohibited.

REF KBI1020

 **96 tests**



KRISHGEN BioSystems |

For US / Europe: toll free +1(888)-970-0827 tel: +1(562)-568-5005
For Asia / India: tel: +91(22)-49198700
Email: sales@krishgen.com

Introduction:


Natalizumab (TYSABRI™) is a humanized monoclonal antibody against the cell adhesion molecule $\alpha 4$ -integrin. Natalizumab is used in the treatment of multiple sclerosis and Crohn's disease. It is co-marketed by Biogen and Élan as Tysabri, and was previously named Antegren. Natalizumab is administered by intravenous infusion every 28 days. The drug is believed to work by reducing the ability of inflammatory immune cells to attach to and pass through the cell layers lining the intestines and blood–brain barrier. Natalizumab has proven effective in treating the symptoms of both diseases, preventing relapse, vision loss, cognitive decline and significantly improving quality of life in people with multiple sclerosis, as well as increasing rates of remission and preventing relapse in Crohn's disease.

Intended Use:

The KRIBIOLISA™ Natalizumab (TYSABRI™) ELISA is used as an analytical tool for quantitative determination of Natalizumab in human serum and plasma.

Principle:

The method employs the quantitative sandwich enzyme immunoassay technique. Antibodies to Natalizumab are pre-coated onto microwells. Samples and standards are pipetted into microwells and human Natalizumab present in the sample are bound by the capture antibody. Then, a HRP (horseradish peroxidase) conjugated anti-Natalizumab antibody is pipetted and incubated. After washing microwells in order to remove any non-specific binding, the ready to use substrate solution (TMB) is added to microwells and color develops proportionally to the amount of Natalizumab in the sample. Color development is then stopped by addition of stop solution. Absorbance is measured at 450 nm.

**Materials Provided:**

| Part | Description | Qty |
|---|--|--------------|
| Anti- Natalizumab Coated Microtiter Plate | 96 well polystyrene microplate (12 strips of 8 wells) coated with Anti-Natalizumab monoclonal antibody. | 1 x 96 wells |
| Natalizumab Standard | Recombinant Natalizumab in a buffered protein base with protein stabilizer and preservatives 0.02% methylisothiazolone and 0.02% bromonitrodioxane.- (lyophilized, concentrated 1 ug/ml) | 2 vials |
| Anti- Natalizumab:HRP Conjugate | Anti- Natalizumab conjugated to Horseradish Peroxidase with protein stabilizer and preservatives 0.02% methylisothiazolone and 0.02% bromonitrodioxane. | 12 ml |
| (1X) Sample Diluent | Buffered protein base with protein stabilizer and preservatives 0.02% methylisothiazolone and 0.02% bromonitrodioxane. | 2 x 50 ml |
| (1X) Standard Diluent | Buffered protein base with protein stabilizer and preservatives 0.02% methylisothiazolone and 0.02% bromonitrodioxane with 1:1000 dilution of human serum | 10 ml |
| (20X) Wash Buffer | 20-fold concentrated solution of buffered surfactant with preservative thiomersol < 0.01%. May turn yellow over time. | 25 ml |
| TMB Substrate | Stabilized Chromogen | 12 ml |
| Stop Solution | 2N Sulfuric Acid | 12 ml |
| Instruction Manual | | 1 no |

Materials to be provided by the End-User:

1. Microtiter Plate Reader able to measure absorbance at 450 nm.
2. Adjustable pipettes and multichannel pipettor to measure volumes ranging from 25 ul to 1000 ul
3. Deionized (DI) water
4. Wash bottle or automated microplate washer
5. Standard graph paper or software for data analysis
6. Timer
7. Absorbent Paper

Handling/Storage:

1. All reagents should be stored at 2°C to 8°C for stability.
2. All the reagents and wash solutions should be used within 12 months from manufacturing date.
3. Before using, bring all components to room temperature (18-25°C). Upon assay completion ensure all components of the kit are returned to appropriate storage conditions.
4. The Substrate is light-sensitive and should be protected from direct sunlight or UV sources.

Health Hazard Warnings:

1. Reagents that contain preservatives may be harmful if ingested, inhaled or absorbed through the skin.
2. For Research Use Only.

**Sample Preparation and Storage:**

Blood is taken by venipuncture. Serum is separated after clotting by centrifugation. Plasma can be used, too. Lipaemic, hemolytic or contaminated samples should not be run. Repeated freezing and thawing should be avoided. If samples are to be used for several assays, initially aliquot samples and keep at -20°C.

Preparation before Use:

Allow samples to reach room temperature prior to assay. Take care to agitate patient samples gently in order to ensure homogeneity.

Test Sample preparation - Serum and Plasma samples have to be diluted 1:1000 (v/v), e.g. for 1:1000 (1 ul sample + 999 ul **(1X) Sample Diluent**) prior to assay. The samples may be kept at 2 - 8°C for up to three days. Long-term storage requires the samples to be kept at -20°C.

Reagent Preparation (all reagents should be diluted immediately prior to use):

1. Label any aliquots made with the kit Lot No and Expiration date and store it at appropriate conditions mentioned.
2. Bring all reagents to Room Temperature before use.
3. To make **Wash Buffer (1X)**; dilute **25 ml of 20X Wash Buffer in 475 ml of DI water**.
4. **Standards Preparation:** Reconstitute the concentrated Standard lyophilized vial with 1 ml of Standard Diluent to obtain a concentration of 1ug/ml. Keep the vial for 15 mins with gentle agitation before making further dilutions. Dilute 640 ul of reconstituted **Standard (1 ug/ml)** with 360 ul of Standard Diluent to generate a **640 ng/ml Standard Solution**. Prepare further **Standards** by serially diluting the Standard Solution as per the below table. Use the Standard Diluent as the Zero Standard (Standard No.0).

| Standard Concentration | Standard Vial | Dilution Particulars |
|------------------------|----------------------|--|
| 1 ug/ml | Lyophilized Standard | Lyophilized Standard provided in the Kit + 1ml of Standard Diluent |
| 640 ng/ml | Standard No.7 | 640ul Reconstituted Standard (1 ug/ml) + 360 ul Standard Diluent |
| 320 ng/ml | Standard No.6 | 500 ul Standard No.7 + 500 ul Standard Diluent |
| 160 ng/ml | Standard No.5 | 500 ul Standard No.6 + 500 ul Standard Diluent |

| Standard Concentration | Standard Vial | Dilution Particulars |
|------------------------|----------------|--|
| 80 ng/ml | Standard No.4 | 500 ul Standard No.5 + 500 ul Standard Diluent |
| 40 ng/ml | Standard No.3 | 500 ul Standard No.4 + 500 ul Standard Diluent |
| 20 ng/ml | Standard No.2 | 500 ul Standard No.3 + 500 ul Standard Diluent |
| 10 ng/ml | Standard No. 1 | 500 ul Standard No.2 + 500 ul Standard Diluent |
| 0 ng/ml | Standard No.0 | Only Standard Diluent |

Procedural Notes:

1. In order to achieve good assay reproducibility and sensitivity, proper washing of the plates to remove excess un-reacted reagents is essential.
2. High Dose Hook Effect may be observed in samples with very high concentrations of Natalizumab. High Dose Hook Effect is due to excess of antibody for very high concentrations of Natalizumab present in the sample.
3. Avoid assay of Samples containing sodium azide (NaN₃), as it could destroy the HRP activity resulting in under-estimation of the amount of Natalizumab.
4. It is recommended that all Standards and Samples be assayed in duplicates.
5. Maintain a repetitive timing sequence from well to well for all the steps to ensure that the incubation timings are same for each well.
6. If the Substrate has a distinct blue color prior to use it may have been contaminated and use of such substrate can lead to compromisation of the sensitivity of the assay.
7. The plates should be read within 30 minutes after adding the Stop Solution.
8. Make a work list in order to identify the location of Standards and Samples.

Assay Procedure:

1. It is strongly recommended that all Standards and Samples be run in duplicates or triplicates. A standard curve is required for each assay. All steps must be performed at 37°C
2. Add **100 ul of prepared Standards or diluted Samples** into the respective wells.
3. Cover the plate and incubate for 60 minutes at 37°C
4. Aspirate and wash plate 4 times with **Wash Buffer (1X)** and blot residual buffer by firmly tapping plate upside down on absorbent paper. Wipe of any liquid from the bottom outside of the microtiter wells as any residue can interfere in the reading step.
5. Add **100 ul of Anti-Natalizumab:HRP Conjugate** into each well.
6. Cover the plate and incubate for 60 minutes at 37°C
7. Aspirate and wash plate 4 times with **Wash Buffer (1X)** and blot residual buffer by firmly tapping plate upside down on absorbent paper. Wipe of any liquid from the bottom outside of the microtiter wells as any residue can interfere in the reading step.
8. Add **100 ul of TMB Substrate** in each well.
9. Incubate the plate at 37°C for 30 minutes in dark. DO NOT SHAKE or else it may result in higher backgrounds and worse precision. Positive wells should turn bluish in color.
10. Pipette out **100 ul of Stop Solution**. Wells should turn from blue to yellow in color.
11. Read the absorbance at 450 nm with a microplate reader.

Calculation of Results:

Determine the Mean Absorbance for each set of duplicate or triplicate Standards and Samples.

Using Standard graph paper, plot the average value (absorbance 450nm) of each standard on the Y-axis versus the corresponding concentration of the standards on the X-axis. Draw the best fit curve through the standard points. To determine the unknown Natalizumab concentrations, find the unknown's Mean Absorbance value on the Y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the X-axis and read the Natalizumab Concentration. If samples were diluted, multiply by the appropriate dilution factor. Software which is able to generate a cubic spline curve-fit or 4PL (2nd order) is best recommended for automated results.


Note:

It is recommended to repeat the assay at a different dilution factor in the following cases:

- If the sample absorbance value is below the first standard.
- If the absorbance value is equivalent or higher than the 640 ng/ml standard.

Typical Data

| Standard Concentration (ng/ml) | Abs A | Abs B | Mean Abs | Interpolated Concentration | % Interpolated Concentration against Actual Concentration |
|--------------------------------|-------|-------|----------|----------------------------|---|
| 0 | 0.073 | 0.065 | 0.069 | --- | --- |
| 10 | 0.099 | 0.109 | 0.104 | 11.9 | 118.8 |
| 20 | 0.146 | 0.111 | 0.128 | 17.7 | 88.4 |
| 40 | 0.298 | 0.261 | 0.279 | 42.3 | 105.8 |
| 80 | 0.666 | 0.523 | 0.595 | 80.4 | 100.5 |
| 160 | 1.397 | 1.125 | 1.261 | 158.0 | 98.7 |
| 320 | 2.365 | 2.009 | 2.187 | 323.6 | 101.1 |
| 640 | 2.818 | 2.844 | 2.831 | 635.3 | 99.3 |

Typical Graph

Abs = absorbance at 450nm

Quality Control:

It is recommended that for each laboratory assay appropriate quality control samples in each run to be used to ensure that all reagents and procedures are correct.

Performance Characteristics of the Kit:

This kit has been validated as per EMA/FDA guidelines in line with ICH Code for Harmonization of Biological Assays.

Sensitivity:

Limit of Quantification: It is defined as the lowest concentration of an analyte that can be determined with an acceptable repeatability and the LOQ was found to be 8.30 ng/ml.

Limit Of Detection: It is defined as the lowest detectable concentration corresponding to a signal of Mean of '0' standard plus 2* SD.

10 replicates of '0' standards were evaluated and the LOD was 7.75 ng/ml.

Specificity:

The antibodies used in the kit are monoclonal antibodies, anti-idiotypic and specific for Natalizumab. The calibrators / standards used are calibrated against commercially sourced biosimilar research grade Natalizumab (TYSABRI).

Precision:

Precision is defined as the percent coefficient of variation (%CV) i.e. standard deviation divided by the mean and multiplied by 100. Assay precision was determined by both intra (n=5 assays) and inter assay (n=5 assays) reproducibility on two pools with low (10 ng/ml), medium (80 ng/ml) and high (640 ng/ml) concentrations. While actual precision may vary from laboratory to laboratory and technician to technician, it is recommended that all operators achieve precision below these design goals before reporting results.

| Pool | Intra Assay %CV | Inter Assay %CV |
|--------|-----------------|-----------------|
| Low | <12% | <12% |
| Medium | <10% | <10% |
| High | <5% | <5% |

Safety Precautions:

- This kit is **For Research Use only**. Follow the working instructions carefully.
- The expiration dates stated on the kit are to be observed. The same relates to the stability stated for reagents
- Do not use or mix reagents from different lots.
- Do not use reagents from other manufacturers.
- Avoid time shift during pipetting of reagents.
- All reagents should be kept in the original shipping container.
- Some of the reagents contain small amount of sodium azide (< 0.1 % w/w) as preservative. They must not be swallowed or allowed to come into contact with skin or mucosa.
- Source materials maybe derived from human body fluids or organs used in the preparation of this kit were tested and found negative for HBsAg and HIV as well as for HCV antibodies. However, no known test guarantees the absence of such viral agents. Therefore, handle all components and all patient samples as if potentially hazardous.
- Since the kit contains potentially hazardous materials, the following precautions should be observed
 - Do not smoke, eat or drink while handling kit material
 - Always use protective gloves
 - Never pipette material by mouth
 - Wipe up spills promptly, washing the affected surface thoroughly with a decontaminant.
- In any case GLP should be applied with all general and individual regulations to the use of this kit.

**References:**

Development and validation of immunoassays to quantify the half-antibody exchange of an IgG4 antibody, natalizumab (Tysabri®) with endogenous IgG4

RI Shapiro, T Plavina, BR Schlain, RB Pepinsky... - ... of pharmaceutical and ..., 2011 - Elsevier

Measurement of serum levels of natalizumab, an immunoglobulin G4 therapeutic monoclonal antibody

T Rispens, A van Leeuwen, A Vennegoor... - Analytical ..., 2011 - Elsevier

Characterization of anti-natalizumab antibodies in multiple sclerosis patients

M Lundkvist, E Engdahl, C Holmén... - Multiple Sclerosis ..., 2013 - journals.sagepub.com

JC-virus seroconversion in multiple sclerosis patients receiving natalizumab

O Outteryck, H Zéphir, J Salleron... - Multiple Sclerosis ..., 2014 - journals.sagepub.com

Occurrence of antibodies against natalizumab in relapsing multiple sclerosis patients treated with natalizumab

PS Sørensen, PE Hyldgaard Jensen... - Multiple Sclerosis ..., 2011 - journals.sagepub.com

Clinical relevance of serum natalizumab concentration and anti-natalizumab antibodies in multiple sclerosis

A Vennegoor, T Rispens, EMM Strijbis... - Multiple Sclerosis ..., 2013 - journals.sagepub.com

A safety and pharmacokinetic study of intravenous natalizumab in patients with MS

WA Sheremata, TL Vollmer, LA Stone... - Neurology, 1999 - AAN Enterprises

SCHEMATIC ASSAY PROCEDURE



1. Remove all components, 30 minutes before adding into the assay plate.




2. Avoid repeated cool-thaw of the components as there will be a loss of activity and this can affect the results.






3.  Pipette **100 ul prepared Standards / diluted Samples** into the respective wells.

4. Cover plate  and incubate for  at 37°C.



5.  Aspirate and wash wells 4 times with **Wash Buffer (1X)**.

6.  Pipette **100 ul Anti- Natalizumab:HRP** into each well.

7. Cover plate  and incubate for  at 37°C

8.  Aspirate and wash wells 4 times with **Wash Buffer (1X)**.

9.  Pipette **100 ul TMB Substrate** into each well.

10. Cover plate  and incubate for  at 37°C.

11.  Pipette **100 ul Stop Solution** into each well.

12. Read absorbance at 450nm with a  microplate reader within  of stopping reaction.

LIMITED WARRANTY

Krishgen Biosystems does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the Products; against defects in products or components not manufactured by Krishgen Biosystems, or against damages resulting from such non-Krishgen Biosystems made products or components. Krishgen Biosystems passes on to customer the warranty it received (if any) from the maker thereof of such non Krishgen made products or components. This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Krishgen Biosystems.

THIS WARRANTY IS EXCLUSIVE. The sole and exclusive obligation of Krishgen Biosystems shall be to repair or replace the defective Products in the manner and for the period provided above. Krishgen Biosystems shall not have any other obligation with respect to the Products or any part thereof, whether based on contract, tort, and strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall Krishgen Biosystems be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of Krishgen Biosystems with respect to the Products. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

Krishgen Biosystems. 2023













THANK YOU FOR USING KRISHGEN PRODUCT!

KRISHGEN BIOSYSTEMS®, GENLISA®, DHARMAPLEX™, GENBULK™, GENLISA™, KRISHZYME®, KRISHGEN®, KRIBIOLISA®, KRISHPLEX®, TITANIUM®, QUALICHEK® are registered trademarks of KRISHGEN BIOSYSTEMS. ©KRISHGEN BIOSYSTEMS. ALL RIGHTS RESERVED.

KRISHGEN BIOSYSTEMS | OUR REAGENTS | YOUR RESEARCH |

TYSABRI™ is the registered trade mark of Biogen Inc., USA.

SYMBOLS KEY

| | |
|---|---|
|  | Anti-Natalizumab Coated Microtiter Plate (12x8 wells) |
|  | Natalizumab Standard |
|  | Conjugate Horseradish Peroxidase |
|  | (1X) Sample Diluent |
|  | (1X) Standard Diluent |
|  | (20X) Wash Buffer |
|  | TMB Substrate |
|  | Stop Solution |
|  | Consult Instructions for Use |
|  | Catalog Number |
|  | Expiration Date |
|  | Storage Temperature |