






# KRIBIOLISA™ Liraglutide (Victoza/Saxenda) ELISA

**REF** : KBI5020

Ver 5.1

**RUO**

Enzyme Immunoassay for the Quantitative Estimation of Liraglutide in human serum and plasma.

<b>RUO</b>	For Research Use Only	<b>REF</b>	Catalog Number
	Store At	<b>LOT</b>	Batch Code
	Manufactured By		Biological Risk
	Expiry Date		Consult Operating Instructions

*For Research Use Only. Purchase does not include or carry the right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of KRISHGEN BioSystems is strictly prohibited.*

KBI5020 **REF**

96 tests 

**KRISHGEN BioSystems** For US/Europe Customers: toll free +1(888)-970-0827 | tel +1(562)-568-5005  
For Asia/India Customers: tel +91(22)-49198700  
Email: sales@krishgen.com | http://www.krishgen.com

**Introduction:**

Liraglutide (NN2211) is a derivative of a human incretin (metabolic hormone), glucagon-like peptide-1 (GLP-1) that is used as a long-acting glucagon-like peptide-1 receptor agonist, binding to the same receptors as does the endogenous metabolic hormone GLP-1 that stimulates insulin secretion. Marketed under the brand name Victoza, it is an injectable drug developed by Novo Nordisk for the treatment of type 2 diabetes.


In 2015, Novo Nordisk began marketing a separate strength in the U.S. and E.U. under the brand name Saxenda as a treatment for adults who are obese or overweight with at least one weight-related comorbid condition.

**Intended Use:**

The KRIBIOLISA™ Liraglutide (Victoza/Saxenda)ELISA is used as for the quantitative determination of Liraglutide in human serum and plasma.

**Principle:**

The method employs a sandwich immunoassay for the determination of Liraglutide. The anti-Liraglutide Antibodies are coated on microtiter plate. Liraglutide standard and Liraglutide present in the samples will bind to coating antibody. Anti-GLP-1 antibody conjugated to HRP is then added which produces a soluble colored product after addition of TMB substrate. The enzyme reaction is stopped by dispensing of stop solution into the wells. The optical density (OD) of the solution at 450 nm is directly proportional to the amount of bound Liraglutide present in the standards or samples.



**PRINCIPLE OF THE KRIBIOLISA™ LIRAGLUTIDE ELISA**

**Materials Provided:**

Part	Description	Qty
Anti-Liraglutide Coated Microtitre Plate	96 well polystyrene microplate (12 strips of 8 wells) coated with Anti-Liraglutide antibody.	1 x 96 wells
Liraglutide Standard	Lyophilized Liraglutide Standard (concentrated – 3000 ng/ml)	2 vials
Anti-GLP-1:HRP Conjugate	Anti-GLP-1:HRP Conjugate prepared in buffer with protein stabilizer and preservatives 0.02% methylisothiazolone and 0.02% bromonitrodioxane.	12 ml
(1X) Sample Diluent	Buffered protein base with protein stabilizer and preservatives 0.02% methylisothiazolone and 0.02% bromonitrodioxane.	2 x 50 ml
(1X) Standard Diluent	Buffered protein base with protein stabilizer and preservatives 0.02% methylisothiazolone and 0.02% bromonitrodioxane with 1:1000 dilution normal human serum	10 ml
(20X) Wash Buffer	20-fold concentrated solution of buffered surfactant with preservative thiomersol < 0.01%. May turn yellow over time.	25 ml
TMB Substrate	Stabilized Chromogen	12 ml
Stop Solution	0.73M Phosphoric Acid	12 ml
Instruction Manual		1 no

**Materials to be provided by the End-User:**

1. Microtiter Plate Reader able to measure absorbance at 450 nm.
2. Adjustable pipettes and multichannel pipettor to measure volumes ranging from 25 ul to 1000 ul
3. Deionized (DI) water
4. Wash bottle or automated microplate washer
5. Graph paper or software for data analysis
6. Timer
7. Absorbent Paper

**Handling/Storage:**

1. Store main kit components at recommended storage temperature indicated on the component label.
2. Before using, bring all components to room temperature (18-25°C). Upon assay completion return all components to appropriate storage conditions.
3. The Substrate is light-sensitive and should be protected from direct sunlight or UV sources.

**Health Hazard Warnings:**

1. Reagents that contain preservatives may be harmful if ingested, inhaled or absorbed through the skin.

**Sample Preparation and Storage:**

Specimens should be clear and non-hemolyzed. Samples should be run at a number of dilutions to ensure accurate quantitation.

Blood is taken by venipuncture. Serum is separated after clotting by centrifugation. Repeated freezing and thawing should be avoided. If samples are to be used for several assays, initially aliquot samples and keep at -20°C.

Samples should be diluted 1:1000 (v/v) for optimal recovery, (for example 1 ul sample + 999 ul sample diluent) prior to assay. In cases where matrix interferences is under or over observed, the samples may be diluted with Sample Diluent accordingly.

The samples may be kept at 2 - 8°C for up to three days. For long-term storage please store at -20°C.

**Note:** Grossly hemolyzed samples are not suitable for use in this assay

**Reagent Preparation (all reagents should be diluted immediately prior to use):**

1. Label any aliquots made with the kit Lot No and Expiration date and store it at appropriate conditions mentioned.
2. Bring all reagents to Room temperature before use.
3. To make **Wash Buffer (1X)**; dilute **25 ml** of **20X Wash Buffer** in **475 ml** of **DI water**.
4. **Standards Preparation:** Reconstitute the concentrated Standard lyophilized vial with 1 ml of Standard Diluent to obtain a concentration of 3000 ng/ml. Keep the vial for 15 mins with gentle agitation before making further dilutions. Dilute 853.3 ul of reconstituted original **Standard (3000 ng/ml)** with 146.7 ul of Standard Diluent to generate a **2560 ng/ml Standard Solution**. Prepare further **Standards** by serially diluting the Standard Solution as per the below table. Use the Standard Diluent as the Zero Standard (Standard No.0).

Standard Concentration	Standard Vial	Dilution Particulars
3000 ng/ml	Original Standard	Original Standard provided in the Kit + 1 ml Standard Diluent (1X)
2560 ng/ml	Standard No.7	853.3 ul Original Standard (3000 ng/ml) + 146.7 ul Standard Diluent (1X)
1280 ng/ml	Standard No.6	500 ul Standard No.7 + 500 ul Standard Diluent (1X)
640 ng/ml	Standard No.5	500 ul Standard No.6 + 500 ul Standard Diluent (1X)
320 ng/ml	Standard No.4	500 ul Standard No.5 + 500 ul Standard Diluent (1X)
160 ng/ml	Standard No.3	500 ul Standard No.4 + 500 ul Standard Diluent (1X)
80 ng/ml	Standard No.2	500 ul Standard No.3 + 500 ul Standard Diluent (1X)
40 ng/ml	Standard No.1	500 ul Standard No.2 + 500 ul Standard Diluent (1X)
0 ng/ml	Standard No. 0	Only Standard Diluent (1X)

Use the Standards as soon as possible upon reconstitution. Discard balance standard after use.

**Procedural Notes:**

1. In order to achieve good assay reproducibility and sensitivity, proper washing of the plates to remove excess un-reacted reagents is essential.
2. Avoid assay of Samples containing sodium azide ( $\text{NaN}_3$ ), as it could destroy the HRP activity resulting in under-estimation of the amount of Liraglutide.
3. It is recommended that the Standards and Samples be assayed in duplicates.
4. Maintain a repetitive timing sequence from well to well for all the steps to ensure that the incubation timings are same for each well.
5. If the Substrate has a distinct blue color prior to use it may have been contaminated and use of such substrate can lead to compromisation of the sensitivity of the assay.
6. The plates should be read within 30 minutes after adding the Stop Solution.
7. Make a work list in order to identify the location of Standards and Samples.

**Assay Procedure:**

1. Pipette **100 ul** of **Standards** and **Samples** to the respective wells.
2. Seal plate and incubate for 1 hour at 37°C.
3. Wash plate 4 times with **Wash Buffer (1X)** and blot residual buffer by firmly tapping plate upside down on absorbent paper. Wipe of any liquid from the bottom outside of the microtiter wells as any residue can interfere in the reading step. All the washes should be performed similarly.
4. Add **100 ul** of **Anti-GLP-1:HRP Conjugate** to each well.
5. Seal plate and incubate for 1 hour at 37°C.

6. Wash plate 4 times with **Wash Buffer (1X)** and blot residual buffer by firmly tapping plate upside down on absorbent paper. Wipe of any liquid from the bottom outside of the microtiter wells as any residue can interfere in the reading step. All the washes should be performed similarly.
7. Pipette **100 ul of TMB Substrate solution** in all wells.
8. Incubate in the dark for 30 minutes at 37°C.
9. Stop reaction by adding **100 ul of Stop Solution** to each well.
10. Read absorbance at 450 nm within 30 minutes of stopping reaction.

**Calculation of Results:**

Determine the Mean Absorbance for each set of duplicate Standards and Samples. Using standard graph paper, plot the average value (absorbance 450nm) of each standard on the Y-axis versus the corresponding concentration of the standards on the X-axis. Draw the best fit curve through the standard points. To determine the unknown Liraglutide concentrations, find the unknown's Mean Absorbance value on the Y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the X-axis and read the concentration. If samples were diluted, multiply by the appropriate dilution factor.

Software which is able to generate a polynomial regression (2<sup>nd</sup> order) or a cubic spline curve-fit is best recommended for automated results.


**Note:**

It is recommended to repeat the assay at a different dilution factor in the following cases:  
 - If the sample absorbance value is below the first standard.

**Typical Data**

Standard Concentration (ng/ml)	Abs A	Abs B	Mean Abs	Interpolated Concentration	% Interpolated Concentration against Actual Concentration
0	0.107	0.127	0.117	--	--
40	0.194	0.160	0.177	33.9	84.6
80	0.224	0.224	0.224	71.0	88.7
160	0.398	0.245	0.321	156.8	98.0
320	0.623	0.450	0.536	372.1	116.3
640	0.862	0.721	0.791	654.5	102.3
1280	1.257	1.255	1.256	1217.0	95.1
2560	2.485	2.062	2.274	2583.0	100.9

**Typical Graph**



Abs = absorbance at 450nm

**Quality Control:**

It is recommended that for each laboratory assay appropriate quality control samples in each run to be used to ensure that all reagents and procedures are correct.

**Performance Characteristics of the Kit:**

This kit has been validated as per EMA/FDA guidelines in line with ICH Code for Harmonization of Biological Assays.

**Sensitivity:**

**Limit Of Detection:** It is defined as the lowest detectable concentration corresponding to a signal of Mean of '0' standard plus 2\* SD. 10 replicates of '0' standards were evaluated and the LOD was found to be less than 40 ng/ml

**Specificity:**

The antibodies used in the kit are monoclonal for high specificity. The standards have been calibrated against commercially sourced Victoza® Injection manufactured by Novo Nordisk.

**Linearity:**

Standards provided in the kit were used for measuring the linearity range of Liraglutide present in matrix. The Detection range provided is 0 - 2560 ng/ml.

**Precision:**

Precision:

Precision is defined as the percent coefficient of variation (%CV) i.e. standard deviation divided by the mean and multiplied by 100. Assay precision was determined by both intra (n=5 assays) and inter assay (n=5 assays) reproducibility on two pools with low (40 ng/ml), medium (128 ng/ml) and high (2560 ng/ml) concentrations. While actual precision may vary from laboratory to laboratory and technician to technician, it is recommended that all operators achieve precision below these design goals before reporting results.

Pool	Intra Assay %CV	Inter Assay %CV
Low	<15%	<15%
Medium	<12%	<12%
High	<12%	<12%

**Safety Precautions:**

- **This kit is For Research Use Only.** Follow the working instructions carefully.
- The expiration dates stated on the kit are to be observed. The same relates to the stability stated for reagents
- Do not use or mix reagents from different lots.
- Do not use reagents from other manufacturers.
- Avoid time shift during pipetting of reagents.
- All reagents should be kept in the original shipping container.
- Some of the reagents contain small amount of sodium azide (< 0.1 % w/w) as preservative. They must not be swallowed or allowed to come into contact with skin or mucosa.
- Since the kit contains potentially hazardous materials, the following precautions should be observed
  - Do not smoke, eat or drink while handling kit material
  - Always use protective gloves
  - Never pipette material by mouth
  - Wipe up spills promptly, washing the affected surface thoroughly with a decontaminant.
- In any case GLP should be applied with all general and individual regulations to the use of this kit.




SCHMATIC ASSAY PROCEDURE

1. Remove all components, 30 minutes before adding into the assay plate.



2. Avoid repeated cool-thaw of the components as there will be a loss of activity and this can affect the results.



3. Pipette **100 ul Standards and diluted Samples** into the respective wells.

4. Cover plate and **incubate** for at 37°C.

5. Aspirate and wash wells 4 times with **Wash Buffer (1X)**.

6. Pipette **100 ul Anti-GLP-1:HRP Conjugate** into each well.

7. Cover plate and **incubate** for at 37°C.

8. Aspirate and wash wells 4 times with **(1X) Wash Buffer**.

9. Pipette **100 ul TMB Substrate** into each well.

10. Cover plate and **incubate** for at 37°C.

11. Pipette **100 ul Stop Solution** into each well.

12. Read absorbance at 450nm with a microplate reader within of stopping reaction.

### Typical Example of a Work List

Well #	Contents	Absorbance at 450nm	Mean Absorbance	Interpolated Conc (ng/ml)
1A	0 Standard			
2A	0 Standard			
1B	40 ng/ml			
2B	40 ng/ml			
1C	80 ng/ml			
2C	80 ng/ml			
1D	160 ng/ml			
2D	160 ng/ml			
1E	320 ng/ml			
2E	320 ng/ml			
1F	640 ng/ml			
2F	640 ng/ml			
1G	1280 ng/ml			
2G	1280 ng/ml			
1H	2560 ng/ml			
2H	2560 ng/ml			
3A	<i>Sample</i>			
4A				
3B	<i>Sample</i>			
4B				

### LIMITED WARRANTY

Krishgen Biosystems does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the Products; against defects in products or components not manufactured by Krishgen Biosystems, or against damages resulting from such non-Krishgen Biosystems made products or components. Krishgen Biosystems passes on to customer the warranty it received (if any) from the maker thereof of such non Krishgen made products or components.

This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Krishgen Biosystems.

THIS WARRANTY IS EXCLUSIVE. The sole and exclusive obligation of Krishgen Biosystems shall be to repair or replace the defective Products in the manner and for the period provided above. Krishgen Biosystems shall not have any other obligation with respect to the Products or any part thereof, whether based on contract, tort, and strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall Krishgen Biosystems be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of Krishgen Biosystems with respect to the Products. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.













Krishgen Biosystems. 2021

### THANK YOU FOR USING KRISHGEN PRODUCT!

KRISHGEN BIOSYSTEMS®, GENLISA®, DHARMAPLEX™, GENBULK™, GENLISA™, KRISHZYME®, KRISHGEN®, KRIBIOLISA®, KRISHPLEX®, TITANIUM®, QUALICHEK® are registered trademarks of KRISHGEN BIOSYSTEMS. ©KRISHGEN BIOSYSTEMS. ALL RIGHTS RESERVED.



KRISHGEN BIOSYSTEMS | OUR REAGENTS | YOUR RESEARCH |

KEY			SYMBOLS
		Anti-Liraglutide Microtiter Plate (12X8 wells)	
		Liraglutide Standard, lyophilized	
29520-0101-0101 HRP-UG-DNJ		Conjugate Horseradish Peroxidase	
		(1X) Sample Diluent	
		(1X) Standard Diluent	
		(20X) Wash Buffer	
		TMB Substrate	
		Stop Solution	
		Consult Instructions for Use	
		Catalogue Number	
		Expiration Date	
		Storage Temperature	