

GENLISA® Mouse Minute Virus Antibody IgG (MVM) ELISA

REF : KLM270

Ver2.4

RUO

Enzyme Immunoassay for the Qualitative Screening of antibodies against Minute Virus Antibody IgG (MVM) in Mouse sera

RUO	For Research Use Only	REF	Catalog Number
	Store At	LOT	Batch Code
	Manufactured By		Biological Risk
	Expiry Date		Consult Operating Instructions

For Use On Research Animals Only. Not for use in diagnostic or therapeutic procedures. Purchase does not include or carry the right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Krishgen Biosystems Private Limited is strictly prohibited

REF KLM270  1 x 96 tests

Krishgen Biosystems Private Limited

For US/Europe Customers: toll free +1(888)-970-0827 | tel +1(562)-568-5005

For Asia/India Customers: +91(22)-49198700

Email: sales1@krishgen.com | <http://www.krishgen.biz> / www.krishgenbio.com

GENLISA® Mouse Minute Virus Antibody IgG (MVM) ELISA

Introduction:

Minute Viruses are enveloped RNA viruses from the Coronaviridae family and part of the Coronavirinae subfamily. With its characteristic surface, the virions appear as a crown like image under the electron microscope and so the viruses are named after the Latin word corona, meaning 'crown' or 'halo'. In animals the viruses infect the respiratory and gastrointestinal systems as well as occasionally affecting the liver and the neurological systems.

The Mouse Minute Virus Antibody IgG (MVM) mainly infect the upper respiratory and gastrointestinal tract. They often result in upper respiratory tract infections (simple colds) in Mouses, causing mild illnesses usually of short lasting nature with a rhinitis, cough, sore throat, as well as fever.

Intended Use:

The GENLISA® Mouse Minute Virus Antibody IgG (MVM) ELISA is used as an analytical tool for qualitative laboratory screening of presence or absence of antibodies against Mouse Minute Virus Antibody IgG (MVM) in the serum of the lab animals.

Principle:

The GENLISA® Mouse Minute Virus Antibody IgG (MVM) Qualitative kit is an Enzyme-Linked Immunosorbent Assay (ELISA). Microtiter plates are precoated with Mouse Minute Virus Antibody IgG (MVM) antigen. Serum is added to microtiter plate. If antibodies against Mouse Minute Virus Antibody IgG (MVM) are present in the sample, it binds to the antigen. Anti-Mouse MVM IgG:HRP conjugate is then added. This immunological reaction results in the formation of an Antigen- Antibody-Enzyme labeled antibody. The wells are washed to remove any unbound reactants as per the wash procedure. TMB Substrate is then added. If any enzyme is present, the Substrate will be hydrolyzed signifying the presence of antibodies against Mouse Minute Virus (MVM) Antibody IgG. This color reaction can be measured spectrophotometrically or observed visually. The color development is proportional to the concentration of antibodies against Mouse Minute Virus Antibody IgG (MVM) in the sample.

Materials Provided:

1. Microtitre coated plate (96 wells) - 1 no
1 holder containing 6 Positive Viral Antigen coated strips ringed in black color and 6 Negative Viral Antigen coated strips ringed in red color.
2. Positive Control – 0.5 ml
3. Negative Control – 0.5 ml
4. Calibrator – 50 ul
5. Anti-Mouse MVM IgG:HRP Conjugate – 5.5 ml
6. (20X) Wash Buffer – 25 ml
7. Dilution Buffer - 30 ml
8. TMB Substrate – 12 ml
9. Stop Solution – 12 ml
10. Instruction Manual

Materials to be provided by the End-User:

1. Microplate Reader able to measure absorbance at 450 nm.
2. Adjustable pipettes to measure volumes ranging from 50 ul to 1000 ul.
3. Deionized (DI) water.
4. Wash bottle or automated microplate washer.
5. Timer.
6. Absorbent paper

GENLISA® Mouse Minute Virus Antibody IgG (MVM) ELISA

Handling/Storage:

1. Reconstitute or dilute only the specific reagents mentioned in the reagent preparation section, when ready to run the assay.
2. Store all kit components at 4°C to 8°C when not in use and do not expose them to temperatures greater than 37°C or less than 2°C.
3. Do not use kit components after the expiration date.
4. Do not repeatedly freeze/thaw the reagents as loss of activity may result.
5. Before using, bring all components to Room Temperature (18-25°C). Upon assay completion return all components to appropriate storage conditions.
6. ELISA plate pouches contain dessicant. Keep the plates sealed in the pouch with dessicant in the refrigerator when not in use.

Health Hazard Warnings:

1. Reagents that contain preservatives may be harmful if ingested, inhaled or absorbed through the skin.
2. Handle Stop Solution carefully. Obtain medical attention in case of accidental ingestion of kit components.
3. Avoid assay of samples containing Sodium azide as it is hazardous.

Reagent Preparation (all reagents should be diluted immediately prior to use):

1. 1X Wash Buffer:
Dilution: To make 1X Wash Buffer, add 25 ml of (20X) Wash Buffer to 475 ml of DI water. This is the working solution.
2. Dilute the calibrator serum 1:51 in dilution buffer. For example: add 5 ul of calibrator serum to 250 ul of Dilution buffer. If not assayed immediately, diluted serum should be stored at -20°C or below.

Specimen Collection and Preparation:

Specimens should be clear and non-hemolyzed. Samples should be run at a number of dilutions to ensure accurate quantitation.

Serum: Use a serum separator tube and allow clotting for 30 minutes, then centrifuge for 10 minutes at 1000 x g. Remove serum layer and assay immediately or store serum samples at <-20°C. Avoid repeated freeze/thaw cycles.

Dilute the serum 1:51 in Dilution Buffer. For example: add 5 ul of serum sample to 250 ul of Dilution Buffer. If not assayed immediately, diluted samples should be stored at -20°C or below.

Procedural Notes:

1. Read all the instructions thoroughly before performing the test.
2. Allow all reagents to reach Room Temperature before beginning and reconstitute or dilute the required reagents.
3. In order to achieve good assay reproducibility and sensitivity, proper washing of the plates to remove excess unreacted reagents is essential.
4. All Controls and Samples should be assayed at least in duplicates.
5. The assay has been optimized to be used with the protocol mentioned. Any deviation from the same may invalidate the results.

Assay Procedure:

1. Bring all reagents to Room Temperature prior to use. It is strongly recommended that all **Controls and Samples** should be run in duplicates or triplicates.
2. The holder contains 12 strips alternating between Negative Viral Antigen ringed in red color and Positive Viral Antigen ringed in black color. Each control, calibrator, and specimen requires two wells: a Negative Viral Antigen coated well and Positive Viral Antigen coated well.

GENLISA® Mouse Minute Virus Antibody IgG (MVM) ELISA

3. Add 50 ul of **Negative Control, Positive Control** and diluted **Sample** to the appropriate wells (Controls are ready to use and do not require any dilution step).
4. Mix the contents in the wells by moving the plate in rapid circular motion, see to it that the contents do not spill. Incubate at Room Temperature for 30 minutes.
5. Aspirate and wash plate 5 times with **Wash Buffer** and blot residual buffer by firmly tapping plate upside down on absorbent paper. Wipe off any liquid from the bottom outside of the microtiter wells as any residue can interfere in the reading step. All the washes should be performed similarly or alternatively a microtiter plate or strip washer may be used.
6. Add **50 ul of Anti-Mouse MVM IgG:HRP Conjugate** to the wells.
7. Incubate at Room Temperature for 30 minutes.
8. Wash the plate as per the instruction given in step 5.
9. Add **100 ul of TMB Substrate** solution into each well of the plate. Incubate for 10 minutes.
10. Add **100 ul of Stop Solution** to each well.
11. Measure the optical density of the wells on a plate reader at 450 nm within 15 minutes.

Interpretation of the Results:

1. It is recommended that each laboratory establish their own criteria for performance of these Research Reagents.
2. In our quality control testing, we use the following criteria:
3. The Positive Control Serum, after subtracting the absorbance in the Negative Control Antigen well, should produce a net absorbance on the Positive Viral Antigen of ≥ 1.00 at 450 nm.
4. A sample may be considered Positive if the index is **1.0 or greater**, if the index is **less than 1.0** then the test is considered as negative.

Formula:

$$\text{Index value} = \text{OD}_{\text{sample/calibrator}} / \text{OD}_{\text{calibrator}}$$

$\text{OD}_{\text{sample/calibrator}}$ = OD of the sampel / control.

$\text{OD}_{\text{calibrator}}$ = OD of calibrator.

Calculation example:

	Positive Antigen well OD	Negative Control Antigen well OD	Differential OD
Negative control	0.12	0.04	0.08
Positive control	1.82	0.03	1.79
Calibrator	0.38	0.10	0.28
Specimen 1	1.10	0.19	0.91
Specimen 2	0.25	0.02	0.23

$$\text{Index value} = \text{OD}_{\text{sample/calibrator}} / \text{OD}_{\text{calibrator}}$$

Differential OD of calibrator	0.28		
Index value of Negative Control	0.08 / 0.28	0.29 index	Valid
Index value of Positive control	1.79 / 0.28	6.39 index	Valid
Index value of specimen 1	0.91 / 0.28	3.25 index	Positive
Index value of specimen 2	0.23 / 0.28	0.82 index	Negative

Expected Values:

The normal value is Negative. Studies have shown that antibodies may take up to 21 days to appear after exposure. Negative specimen results should be reviewed in relation to a possible exposure date. All Positive specimen results should be confirmed by an alternate method.

GENLISA® Mouse Minute Virus Antibody IgG (MVM) ELISA

Precautions:

1. Do not mix reagents from different kits or lots. Reagents and/or antibodies from different manufacturers should not be used with this set.
2. Substrate is light and heat sensitive hence do not expose it to direct sunlight while pipetting or incubating.
3. Samples and kit reagents after use should be disposed off observing appropriate regulations.
4. If necessary it is recommended that the results should be confirmed by an alternative method.
5. Do not dilute or adulterate test reagents or use samples not called for in the test procedure.

LIMITED WARRANTY

Krishgen Biosystems Private Limited does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the Products; against defects in products or components not manufactured by Krishgen Biosystems Private Limited, or against damages resulting from such non-Krishgen Biosystems Private Limited made products or components. Krishgen Biosystems Private Limited passes on to customer the warranty it received (if any) from the maker thereof of such non Krishgen made products or components. This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Krishgen Biosystems Private Limited.

THIS WARRANTY IS EXCLUSIVE. The sole and exclusive obligation of Krishgen Biosystems Private Limited shall be to repair or replace the defective Products in the manner and for the period provided above. Krishgen shall not have any other obligation with respect to the Products or any part thereof, whether based on contract, tort, strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall Krishgen Biosystems Private Limited be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of Krishgen Biosystems Private Limited with respect to the Products. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

Krishgen Biosystems Private Limited. 2025

THANK YOU FOR USING A KRISHGEN PRODUCT!

KRISHGEN BIOSYSTEMS PRIVATE LIMITED®, DHARMAPLEX®, GENBULK®, GENLISA®, KRISHZYME®, KRISHGEN®, KRIBIOLISA®, KRISHPLEX®, TITANIUM®, QUALICHEK® are registered trademarks of KRISHGEN BIOSYSTEMS PRIVATE LIMITED.
©KRISHGEN BIOSYSTEMS PRIVATE LIMITED. ALL RIGHTS RESERVED.

KRISHGEN BIOSYSTEMS PRIVATE LIMITED | OUR REAGENTS | YOUR RESEARCH |

PLATE MAP

Assay Name: _____

Date: _____

Lot No: _____

Plate: _____

	Positive (Black)	Negative (Red)										
	1	2	3	4	5	6	7	8	9	10	11	12
A	Reactive Control		Sample 6		Sample 14		Sample 22		Sample 30		Sample 38	
B	Negative Control		Sample 7		Sample 15		Sample 23		Sample 31		Sample 39	
C	Calibrator		Sample 8		Sample 16		Sample 24		Sample 32		Sample 40	
D	Sample 1		Sample 9		Sample 17		Sample 25		Sample 33		Sample 41	
E	Sample 2		Sample 10		Sample 18		Sample 26		Sample 34		Sample 42	
F	Sample 3		Sample 11		Sample 19		Sample 27		Sample 35		Sample 43	
G	Sample 4		Sample 12		Sample 20		Sample 28		Sample 36		Sample 44	
H	Sample 5		Sample 13		Sample 21		Sample 29		Sample 37		Sample 45	